

**JBW**

# CERTIFICATE OF VALIDATION

AS FOUND DATA

## NIST\*-Traceable Certification for SAS Air Sampler

\*National Institute of Standards and Technology

Model: SUPER " 100 "

Date: 3-12-19

Serial No.: 00-C-01686 w/ss head 4508

Val. Technician: [Signature]

Company: Aerobiology Laboratory

Checked by: [Signature]

Temperature, °F: 70

Barometric Pressure, in. Hg: 30.2

Temp. and Pressure Correction Factor: 1.006

Anemometer Reading, meters/second: .293

Corrected Air Velocity, meters/second: .295

Air Flow, liters/minute: 103.7

Display Setting (Sample Volume)	Sampling Time (Minutes)
100 liters:	1.0 Minutes
500 liters:	5.0 Minutes
1000 liters:	10.0 Minutes

**Comments:**

1. Accuracy =  $\pm 5\%$  Total
2. Validated at battery output 9.9 volts.  
BBL TSA Petri
3. Validated with contact plate in place. Lot 8355946 Exp. 6-14-19
4. Anemometer serial number 59030121 (see attached certificate of calibration).

**JBW & Associates, Inc.**  
10242 Little Rock Lane  
Frederick, Maryland 21702

**Bioscience International**  
Telephone: 301-230-0072