

Certificate of Calibration

AeroTrap™ Rotameter Calibration

Serial Number: B030066

Date Calibrated: 03/07/19

Calibration Due Date: 03/07/20

Flow Calibration

This is to certify that the rotameter listed above has been calibrated using a TSI Primary calibrator listed below which is calibrated to U.S. National Institute of Standards Technology (NIST) traceable standards according to the procedure outlined in TSI 9010471.

AMBIENT CONDITIONS: Temperature $74 \pm 3^{\circ}\text{F}$ Relative Humidity $50 \pm 10\%$

Description	MFR	Model	Serial #	Calibration Due Date
Primary Calibrator	TSI	4046	4046 0540 001	04/2019

QA Approval By: _____



Information contained in this document should not be reproduced in any form without the written consent of Aerobiology Laboratory Associates, Inc. It is for reference only and cannot be used as a form of endorsement by any private or governmental regulatory body.

Aerobiology Laboratory Associates, Inc.
11800 Sunrise Valley Dr., Suite 300
Reston, VA 20191
Phone: 703-648-9150 · Fax: 703-648-3919

 **AEROBIOLOGY LABORATORY**
ASSOCIATES, INCORPORATED
MICROBIOLOGY SPECIALISTS