

**JBW**

# CERTIFICATE OF CALIBRATION

FINAL DATA

## NIST\*-Traceable Certification for SAS Air Sampler

\*National Institute of Standards and Technology

Model: SUPER " 100 "

Date: 9-22-18

Serial No. : 14-D-09351 w/ss head 5600

Next Cal. Due: 9-22-19

Company: Aerobiology Lab. Associates

Cal. Technician: [Signature]

Checked by: [Signature]

Temperature, °F: 70

Barometric Pressure, in. Hg: 30.0

Temp. and Pressure Correction Factor: 1.0

Anemometer Reading, meters/second: .286

Corrected Air Velocity, meters/second: .286

Air Flow, liters/minute: 100.0

Display Setting (Sample Volume)	Sampling Time (Minutes)
100 liters:	1.0 Minutes
500 liters:	5.0 Minutes
1000 liters:	10.0 Minutes

**Comments:**

1. Accuracy =  $\pm 5\%$  Total
2. Calibrated at battery output 9.6 volts. BBL TSA Petri
3. Calibrated with contact plate in place. LOT 8156820 Exp. 11-13-18
4. Anemometer serial number 66050014 (see attached certificate of calibration).

**JBW & Associates, Inc.**  
10242 Little Rock Lane  
Frederick, Maryland 21702

**Bioscience International**  
Telephone: 301-230-0072