

JBW

CERTIFICATE OF CALIBRATION

FINAL DATA

NIST*-Traceable Certification for SAS Air Sampler

*National Institute of Standards and Technology

Model: SUPER " 90 "

Date: 8-22-19

Serial No. : 97/D 22645 w/ss head 8519

Next Cal. Due: 8-22-20

Company: Aerobiology Laboratory

Cal. Technician: _____

ID: Aero #4

Checked by: [Signature]

Temperature, °F: 70

Barometric Pressure, in. Hg: 30.0

Temp. and Pressure Correction Factor: 1.0

Anemometer Reading, meters/second: .257

Corrected Air Velocity, meters/second: .257

Air Flow, liters/minute: 90.0

Display Setting (Sample Volume)	Sampling Time (Minutes)
100 liters:	1.1 Minutes
500 liters:	5.5 Minutes
1000 liters:	11.0 Minutes

Comments:

1. Accuracy = ± 5% Total
2. Calibrated at battery output 9.5 volts.
Hardy TSA Petri
3. Calibrated with contact plate in place. _____

Lot G60-443133 Exp. 10-18-19

4. Anemometer serial number T95451921005 (see attached certificate of calibration).

JBW & Associates, Inc.
10242 Little Rock Lane
Frederick, Maryland 21702

Bioscience International
Telephone: 301-230-0072

JBW

CERTIFICATE OF VALIDATION

AS FOUND DATA

NIST*-Traceable Certification for SAS Air Sampler

*National Institute of Standards and Technology

Model: SUPER " 90 "

Date: 8-22-19

Serial No.: 97/D 22645 w/ss head 8519

Val. Technician: J. B. White

Company: Aerobiology Laboratory
ID: Aero #4

Checked by: [Signature]

Temperature, °F: 70

Barometric Pressure, in. Hg: 30.0

Temp. and Pressure Correction Factor: 1.0

Anemometer Reading, meters/second: .267

Corrected Air Velocity, meters/second: .267

Air Flow, liters/minute: 93.5

Display Setting (Sample Volume)	Sampling Time (Minutes)
100 liters:	1.1 Minutes
500 liters:	5.5 Minutes
1000 liters:	11.0 Minutes

Comments:

1. Accuracy = ± 5% Total
2. Validated at battery output 9.8 volts. Hardy TSA Petri
3. Validated with contact plate in place. Lot G60-443133 Exp. 10-18-19
4. Anemometer serial number T95451921005 (see attached certificate of calibration).

JBW & Associates, Inc.
10242 Little Rock Lane
Frederick, Maryland 21702

Bioscience International
Telephone: 301-230-0072



CERTIFICATE OF CALIBRATION AND TESTING

TSI Incorporated, 500 Cardigan Road, Shoreview, MN 55126 USA
 Tel: 1-800-874-2811 1-651-490-2811 Fax: 1-651-490-3824 <http://www.tsi.com>

ENVIRONMENT CONDITIONS			MODEL	9545
TEMPERATURE	74.4 (23.6)	°F (°C)	SERIAL NUMBER	T95451921005
RELATIVE HUMIDITY	27	%RH		
BAROMETRIC PRESSURE	29.17 (987.8)	inHg (hPa)		

<input checked="" type="checkbox"/> AS LEFT	<input checked="" type="checkbox"/> IN TOLERANCE
<input type="checkbox"/> AS FOUND	<input type="checkbox"/> OUT OF TOLERANCE

- CALIBRATION VERIFICATION RESULTS -

TEMPERATURE VERIFICATION				SYSTEM T-100			Unit: °F (°C)
#	STANDARD	MEASURED	ALLOWABLE RANGE	#	STANDARD	MEASURED	ALLOWABLE RANGE
1	32.0 (0.0)	32.3 (0.2)	31.5~32.5 (-0.3~0.3)	2	140.0 (60.0)	140.2 (60.1)	139.5~140.5 (59.7~60.3)

HUMIDITY VERIFICATION				SYSTEM H-100			Unit: %RH
#	STANDARD	MEASURED	ALLOWABLE RANGE	#	STANDARD	MEASURED	ALLOWABLE RANGE
1	10.0	10.1	7.8~12.2	4	70.0	69.4	67.8~72.2
2	30.0	29.2	27.8~32.2	5	90.0	89.6	87.8~92.2
3	50.0	49.3	47.8~52.2				

VELOCITY VERIFICATION				SYSTEM V-109			Unit: ft/min (m/s)
#	STANDARD	MEASURED	ALLOWABLE RANGE	#	STANDARD	MEASURED	ALLOWABLE RANGE
1	0 (0.00)	0 (0.00)	-3~3 (-0.02~0.02)	7	639 (3.24)	641 (3.25)	620~658 (3.15~3.34)
2	35 (0.18)	35 (0.18)	32~38 (0.16~0.19)	8	984 (5.00)	982 (4.99)	954~1013 (4.85~5.15)
3	65 (0.33)	64 (0.33)	62~68 (0.31~0.34)	9	1466 (7.44)	1469 (7.46)	1422~1509 (7.22~7.67)
4	100 (0.51)	99 (0.50)	97~102 (0.49~0.52)	10	2497 (12.69)	2502 (12.71)	2422~2572 (12.31~13.07)
5	159 (0.81)	160 (0.81)	155~164 (0.78~0.83)	11	4500 (22.86)	4526 (22.99)	4365~4635 (22.18~23.55)
6	327 (1.66)	328 (1.67)	317~337 (1.61~1.71)	12	5800 (29.46)	5798 (29.45)	5626~5974 (28.58~30.35)

TSI does hereby certify that the above described instrument conforms to the original manufacturer's specification (not applicable to As Found data) and has been calibrated using standards whose accuracies are traceable to the United States National Institute of Standards and Technology (NIST) or has been verified with respect to instrumentation whose accuracy is traceable to NIST, or is derived from accepted values of physical constants. TSI's calibration system is registered to ISO-9001:2015.

<table border="0"> <tr> <th>Measurement Variable</th> <th>System ID</th> <th>Last Cal.</th> <th>Cal. Due</th> </tr> <tr> <td>Temperature</td> <td>E003617</td> <td>01-14-19</td> <td>07-31-19</td> </tr> <tr> <td>Humidity</td> <td>E003296</td> <td>08-15-18</td> <td>08-31-19</td> </tr> <tr> <td>Pressure</td> <td>E001557</td> <td>04-17-19</td> <td>10-31-19</td> </tr> <tr> <td>Temperature</td> <td>E001553</td> <td>04-11-19</td> <td>10-31-19</td> </tr> </table>	Measurement Variable	System ID	Last Cal.	Cal. Due	Temperature	E003617	01-14-19	07-31-19	Humidity	E003296	08-15-18	08-31-19	Pressure	E001557	04-17-19	10-31-19	Temperature	E001553	04-11-19	10-31-19	<table border="0"> <tr> <th>Measurement Variable</th> <th>System ID</th> <th>Last Cal.</th> <th>Cal. Due</th> </tr> <tr> <td>Temperature</td> <td>E003305</td> <td>01-14-19</td> <td>07-31-19</td> </tr> <tr> <td>Pressure</td> <td>E010288</td> <td>05-20-19</td> <td>11-30-19</td> </tr> <tr> <td>DC Voltage</td> <td>E005749</td> <td>08-23-18</td> <td>08-31-19</td> </tr> <tr> <td>Velocity</td> <td>E005916</td> <td>08-14-17</td> <td>08-31-20</td> </tr> </table>	Measurement Variable	System ID	Last Cal.	Cal. Due	Temperature	E003305	01-14-19	07-31-19	Pressure	E010288	05-20-19	11-30-19	DC Voltage	E005749	08-23-18	08-31-19	Velocity	E005916	08-14-17	08-31-20
Measurement Variable	System ID	Last Cal.	Cal. Due																																						
Temperature	E003617	01-14-19	07-31-19																																						
Humidity	E003296	08-15-18	08-31-19																																						
Pressure	E001557	04-17-19	10-31-19																																						
Temperature	E001553	04-11-19	10-31-19																																						
Measurement Variable	System ID	Last Cal.	Cal. Due																																						
Temperature	E003305	01-14-19	07-31-19																																						
Pressure	E010288	05-20-19	11-30-19																																						
DC Voltage	E005749	08-23-18	08-31-19																																						
Velocity	E005916	08-14-17	08-31-20																																						

 CALIBRATED

May 21, 2019

 DATE



Calibration complies with ISO/IEC 17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 1015-9767818

Traceable® Certificate of Calibration for Stopwatch

Manufactured for and distributed by: VWR International LLC Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA, 19087

Instrument Identification:

Model: 62379-058, S/N: 181557362 Manufacturer: Control Company

Standards/Equipment:

Table with 4 columns: Description, Serial Number, Due Date, NIST Traceable Reference. Row 1: Non-Contact Frequency Counter, 26.66887, 17 May 2019, 1000425907

Certificate Information:

Technician: 332 Procedure: CAL-01 Cal Date: 30 Aug 2018 Cal Due Date: 30 Aug 2020
Test Conditions: 64.69%RH 23.62°C 1016mBar

Calibration Data: (New Instrument)

Table with 11 columns: Unit(s), Nominal, As Found, In Tol, Nominal, As Left, In Tol, Min, Max, ±U, TUR. Row 1: sec/24hr, N.A., N.A., In Tol, 0.000, -0.100, Y, -0.432, 0.432, 0.041, >4:1

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on last results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min=As Left Nominal(Rounded) - Tolerance; Max= As Left Nominal(Rounded) + Tolerance;

Nicol Rodriguez, Quality Manager

Asron Justice, Technical Manager

Note:

Maintaining Accuracy:

In our opinion once calibrated your Stopwatch should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Stopwatch change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.control3.com

Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2008 Quality Certified by DNV GL, Certificate No. CERT-01805-2008-AQ-HOU-RVA.
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).