

BIOSCIENCE INTERNATIONAL

Innovative Microbiology Products
11333 Woodglen Drive • Rockville, Maryland 20852
301.231.7400 • www.biosci-intl.com • fax: 301.231.7277

CERTIFICATE OF CALIBRATION

Model:	SAS Super 90
Air Sampler Serial #:	97/D22645
Air Sampler Head Serial #:	8519
Customer:	Aerobiology Laboratory
Customer Asset ID #:	Aero #4
Calibration performed at:	JBW site 10242 Little Rock Ln Frederick, MD 21702

Cal. Date:	22 Aug 2020
Cal. Due:	22 Aug 2021 (12 months)
Procedure:	EOP-030
Certificate #:	97/D22645-2065
Volume sampled (L):	1000
Calibrated w/media type:	BBL TSA Petri
Plate Lot #:	0129068
Plate Exp Date:	10/23/2020

	As Found	In Tolerance	As Left	In Tolerance	Acceptable Range
Battery output (Volts):	9.9	n/a	9.6	n/a	>8.2
Temperature (F°):	70.2	n/a	70.2	n/a	59 - 95
Barometric pressure (in. HG):	29.9	n/a	29.9	n/a	n/a
Time to sample 1000 Liters (min)	10.71	n/a	10.92	n/a	n/a
Temp. & Pressure Standardization Factor:	1.00	n/a	1.00	n/a	n/a
Air velocity reading (ft/min)	52.0	n/a	51.0	n/a	n/a
Air velocity reading (m/sec)	0.264	n/a	0.259	n/a	n/a
Standardized air velocity reading (m/sec)	0.264	n/a	0.259	n/a	n/a
Standardized Air Flow (L/min)	93.4	Yes	91.6	Yes	85.5 - 94.5

Additional heads inspected and determined to be within +/-2%: n/a
 Additional service, preventative maintenance, or calibration notes: n/a

Bioscience International certifies that the above described instrument conforms to the original manufacturer's tolerances for the parameters listed (not applicable to As Found data) & has been calibrated in accordance with ISO 17025:2017 guidelines using standards whose accuracies are traceable to the U.S. National Institute of Standards & Technology, have been verified with respect to instrumentation whose accuracy is traceable to NIST, or are derived from accepted values of physical constants. CMC test uncertainty is +/-2.2%. Instruments are calibrated with a test uncertainty ratio of 4:1 or greater whenever possible, with uncertainty defined as within a 95% confidence interval using a coverage factor of k = 2. In all cases, statistical methods are used to minimize uncertainty using the best commercially available methods. In Tolerance conditions are based on test results falling within the Acceptable Range. Measurement uncertainty is provided separately & independent of the decision rule. Voltage readings are for preventative maintenance purposes & not part of the calibration; values other than voltage, temperature, pressure, & air velocity are calculated values. Calibration results relate only to the items listed above; e.g., the instrument should be recalibrated prior to switching to a different media size (e.g., from 90mm Petri dishes to 55mm contact plates or vice versa).

Measurement Standards			
ID	Description	Last Cal.	Cal. Due
J-T95452022004	Velocity	5/29/2020	5/29/2021
J-6530-181788940	Temperature & Pressure	12/7/2018	12/7/2020

Work performed by / date: M. Alt / 8-22-20 Reviewed by / date: James Kay / 8-22-20