

# BIOSCIENCE INTERNATIONAL

Innovative Microbiology Products  
11333 Woodglen Drive • Rockville, Maryland 20852  
301.231.7400 • www.biosci-intl.com • fax: 301.231.7277

## CERTIFICATE OF CALIBRATION

|                            |   |
|----------------------------|---|
| Model:                     | SAS Super 100   |
| Air Sampler Serial #:      | 14-D-09023  |
| Air Sampler Head Serial #: | 5600  |
| Customer:                  | Aerobiology Lab   |
| Customer Asset ID #:       | n/a   |
| Calibration performed at:  | JBW site<br>10242 Little Rock Ln<br>Frederick, MD 21702 |

|                          |                         |
|--------------------------|-------------------------|
| Cal. Date:               | 26 Mar 2022             |
| Cal. Due:                | 26 Mar 2023 (12 months) |
| Procedure:               | EOP-030                 |
| Certificate #:           | 14-D-09023-2646         |
| Volume sampled (L):      | 1000                    |
| Calibrated w/media type: | BBL TSA Petri           |
| Plate Lot #:             | 1302110                 |
| Plate Exp Date:          | 4/14/2022               |

|   | As Found | In Tolerance | As Left | In Tolerance | Acceptable Range |
|---|----------|--------------|---------|--------------|------------------|
| Battery output (Volts):                   | 9.9      | n/a          | 9.6     | n/a          | >8.2             |
| Temperature (F°):                         | 71.5     | n/a          | 71.5    | n/a          | 59 - 95          |
| Barometric pressure (in. HG):             | 29.80    | n/a          | 29.80   | n/a          | n/a              |
| Time to sample 1000 Liters (min)          | 10.37    | n/a          | 10.00   | n/a          | n/a              |
| Temp. & Pressure Standardization Factor:  | 0.99     | n/a          | 0.99    | n/a          | n/a              |
| Air velocity reading (ft/min)             | 54.0     | n/a          | 56.0    | n/a          | n/a              |
| Air velocity reading (m/sec)              | 0.274    | n/a          | 0.284   | n/a          | n/a              |
| Standardized air velocity reading (m/sec) | 0.272    | n/a          | 0.283   | n/a          | n/a              |
| Standardized Air Flow (L/min)             | 96.4     | Yes          | 100.0   | Yes          | 95 - 105         |

Additional heads inspected and determined to be within +/-2%: n/a  
Additional service, preventative maintenance, or calibration notes: n/a

*Bioscience International certifies that the above described instrument conforms to the original manufacturer's tolerances for the parameters listed (not applicable to As Found data) & has been calibrated in accordance with ISO 17025:2017 guidelines using standards whose accuracies are traceable to the U.S. National Institute of Standards & Technology, have been verified with respect to instrumentation whose accuracy is traceable to NIST, or are derived from accepted values of physical constants. CMC test uncertainty is +/-2.2%. Instruments are calibrated with a test uncertainty ratio of 4:1 or greater whenever possible, with uncertainty defined as within a 95% confidence interval using a coverage factor of k = 2. In all cases, statistical methods are used to minimize uncertainty using the best commercially available methods. In Tolerance conditions are based on test results falling within the Acceptable Range. Measurement uncertainty is provided separately & independent of the decision rule. Voltage readings are for preventative maintenance purposes & not part of the calibration; values other than voltage, temperature, pressure, & air velocity are calculated values. Calibration results relate only to the items listed above; e.g., the instrument should be recalibrated prior to switching to a different media size (e.g., from 90mm Petri dishes to 55mm contact plates or vice versa).*

### Measurement Standards

| ID                   | Description            | Last Cal. | Cal. Due  |
|----------------------|------------------------|-----------|-----------|
| J-T95451921005       | Velocity               | 9/1/2021  | 9/1/2022  |
| J-10510922-200515237 | Temperature & Pressure | 8/31/2020 | 8/31/2022 |

Work performed by / date:

*Blaise Bartholm 3-26-22*

Reviewed by / date:

*[Signature] 3-26-22*





Calibration complies with ISO/IEC  
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 6530-11510401

Traceable® Certificate of Calibration for Digital Barometer

Manufactured for and distributed by : VWR International LLC Radnor Corporate Center, Bldg 1, Ste 200, 100 Matsonford Road, Radnor, PA, 19087

Instrument Identification:

Model: 10510-922,

S/N: 200515237

Manufacturer: Control Company

Standards/Equipment:

| Description               | Serial Number | Due Date    | NIST Traceable Reference |
|---------------------------|---------------|-------------|--------------------------|
| Digital Barometer         | D4540Q01      | 01 Nov 2020 | 1000447551               |
| Digital Thermometer       | 130070752     | 10 Mar 2021 | 4000-11170557            |
| Chilled Mirror Hygrometer | 44654/2H3737  | 25 Nov 2021 | 17811                    |
| Climate Chamber           | W613.0046     |             |                          |

Certificate Information:

Technician: 57

Procedure: CAL-31

Cal Date: 31 Aug 2020

Cal Due Date: 31 Aug 2022

Test Conditions: 55.92%RH 24.51°C 1007mBar

Calibration Data: (New Instrument)

| Unit(s) | Nominal | As Found | In Tol | Nominal | As Left | In Tol | Min   | Max   | ±U   | TUR  |
|---------|---------|----------|--------|---------|---------|--------|-------|-------|------|------|
| %RH     | N.A.    | N.A.     |        | 50.97   | 53      | Y      | 48    | 54    | 0.74 | >4:1 |
| °C      | N.A.    | N.A.     |        | 25.22   | 25.0    | Y      | 24.82 | 25.62 | 0.05 | >4:1 |
| mb/hPa  | N.A.    | N.A.     |        | 805.75  | 806     | Y      | 802   | 810   | 0.62 | >4:1 |
| mb/hPa  | N.A.    | N.A.     |        | 909.75  | 911     | Y      | 906   | 914   | 0.62 | >4:1 |
| mb/hPa  | N.A.    | N.A.     |        | 1014.20 | 1015    | Y      | 1010  | 1018  | 0.62 | >4:1 |

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO Guide to the Expression of Uncertainty in Measurement : (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ± U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min=As Left Nominal(Rounded) - Tolerance; Max= As Left Nominal(Rounded) + Tolerance;

*Nicol Rodriguez*

Nicol Rodriguez, Quality Manager

*Marisa Elms*

Marisa Elms, Technical Manager

Note :

Maintaining Accuracy:

In our opinion once calibrated your Digital Barometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Digital Barometer change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

Issue Date : 31 Aug 2020

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598  
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2017 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.  
Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01895-2006-AQ-HOU-ANAB.  
International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).